



ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Saket Nagar, Bhopal (M.P.) – 462020

Website : www.aiimsbhopal.edu.in

CORRIGENDUM & DATE EXTENSION

E Tender No. 022 AIIMS Bhopal/CS/Microbiology Equipments/2018 - 19/01 Dated: 29.06.2018

CORRIGENDUM

A. DARK FIELD MICROSCOPE

SI	Original Specifications	Revised Specifications
1	Should have at least 7 fold coded nosepiece to attach minimum of 7 objectives.	Should have at least 6 fold coded nosepiece to attach minimum of 6 objectives.
2	Should have the following objectives -5x/ NA 0.12, 10x/ NA 0.25, 20x/ NA 0.40; 40x/ NA 0.75, 63x/ NA 0.80; 100x/ NA1.25-0.60 oil	Should have the following objectives -4x OR 5x/ NA 0.12; 10x/ NA 0.25; 20x/ NA 0.40; 40x/ NA 0.75; 60x OR 63x/ NA 0.80; 100x/ NA1.25-0.60 oil
3	Should have Trinocular tube with field of view 25 or better variable beam splitters, 3 beam splitter positions: 100% eyepieces, 50% eyepieces:50% documentation port, 100% documentation port.	Should have Trinocular tube with field of view 25 or better variable beam splitters, 3 beam splitter positions 100% eyepieces, 50% eyepieces:50% documentation port OR 20% eyepieces:80% documentation port. 100% documentation port

B. REAL TIME PCR

SI	Original Specifications	Revised Specifications
1	System should use high intensity LED/Halogen/Xenon as excitation source and CCD camera for accurate fluorescent detection.	System should use high intensity LED/Halogen/Xenon as excitation source and CCD camera/ Photodiode for accurate fluorescent detection.
2	Instrument manufacturer must also be offering US FDA and CE IVD approved kits for quantification of Hepatitis B Virus, Hepatitis C Virus and Human Immunodeficiency Virus and Hepatitis C Virus genotyping	Instrument manufacturer must also be offering documentary evidence of certification showing compatibility of the instrument with US FDA and CE IVD approved kits for quantification of Hepatitis B Virus, Hepatitis C Virus and Human Immunodeficiency Virus and Hepatitis C Virus genotyping

C.- Gel Imaging System:-

SI	Original Specifications	Revised Specifications
1	Image sensor Peltier cooled 16-bit 8.3 megapixel CCD or better.	Image sensor Peltier cooled 16-bit 4-8 megapixel CCD or better.
2	Motorized zoom lens 10mm-100mm. High Efficiency Diopter lens.	Motorized zoom lens

डॉ. एस बालकृष्णन
Dr. S. BALAKRISHNAN
आर्चाय एवं विभागाध्यक्ष / Professor & Head
भेषजगुण विज्ञान विभाग / Department of Pharmacology
एम्स भोपाल, साकेत नगर, भोपाल - 462024 (M.P.)
23.7.18

Dr. Arun Kumar Kapoor
Professor & Head
Dept. of Pathology & Lab Medicine
All India Institute of Medical Sciences
Saket Nagar, Bhopal-462024

Head of the Department
of Microbiology
Gandhi Medical College
Bhopal (M.P.)

Dr. DEBASIS BISWAS
Professor & Head
Department of Microbiology
All India Institute of Medical Sciences
AIIMS, Saket Nagar, Bhopal-462020



ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Saket Nagar, Bhopal (M.P.) – 462020

Website : www.aiimsbhopal.edu.in

3.	Software should have comprehensive modules for i. 1-D electrophoresis gel, Blot analysis. ii. Array and microplate analysis. iii. Colony counting and 2-D spot analysis.	Software should have comprehensive modules for i. 1-D electrophoresis gel, Blot analysis. ii. Array and microplate analysis.
----	---	--

Balakrishnan
23/7/18

डॉ. एस बालकृष्णन
Dr. S. BALAKRISHNAN

आर्चाय एवं विभागाध्यक्ष / Professor & Head
भेषजगुण विज्ञान विभाग / Department of Pharmacology
एम्स भोपाल, साकेत नगर, भोपाल - 462024 (म.प्र.)
AIIMS Bhopal, Saket Nagar, Bhopal - 462024 (M.P.)

Neelekamal Kapoor
23/7/18

Dr. Neelekamal Kapoor
Professor & Head
Deptt. of Pathology & Lab Medicine
All India Institute of Medical Sciences
Saket Nagar, Bhopal-462024

Debasis Biswas
23/7/18

Dr. DEBASIS BISWAS
Professor & Head
Department of Microbiology
All India Institute of Medical Sciences
AIIMS, Saket Nagar, Bhopal-462024

Debasis Biswas
Head of the Department
of Microbiology
Gandhi Medical College
Bhopal (M.P.)



ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Saket Nagar, Bhopal (M.P.) – 462020

Website : www.aiimsbhopal.edu.in

Date: 24/09/2018

1st EXTENSION OF DATE FOR SUBMISSION OF e-TENDER

1. Last date for online Submission of e-Tender No.: 022 Dated: 29.06.2018 for “Equipments for Department of Microbiology, AIIMS Bhopal” is extended to **01st October 2018** Upto 17.00Hrs.
2. Tender will be opened **03rd October 2018** at 15:00 Hrs. in the Central Stores, Hospital Building, 3rd Floor, Saket Nagar, Bhopal.

For & on behalf of the Director
AIIMS Bhopal