

Minutes of Pre-Bid Clarification

The Pre-Bid Meeting was held on 15.12.2015 at 15:00 Hrs. in the office of Medical Superintendent, AIIMS Bhopal Hospital with reference to AIIMS Bhopal E-tender No. AIIMSBPLHOSP151602304122015 for Supply, Installation, Testing & Commissioning of Automatic Tourniquet System in Department of Orthopedics at AIIMS Bhopal Hospital.

No Bidder appeared during the Pre-Bid meeting scheduled as above. However, following Bidder Company forwarded their queries to the email ID of e-Tendering Department of AIIMS Bhopal Hospital at etender@aiimsbhopal.edu.in as follows:-


Sr. No.	Bidder Company Name
1.	M/s APH Medical System

The Pre-Bid Query placed before to the User Department's Subject Technical Expert by the E-Tendering Department for their attention and user Department was discussed queries raised by the bidders in Technical Terms and decided to communicate all interested participating bidders through uploading the same on the E-Tendering Portal/AIIMS Bhopal website and CPP Portal.

The Pre-Bid Queries replies applicable for referenced Tender ID is as under:-

Query No.	Bidder's Query	AIIMS Bhopal reply (whereas applicable in entire Tender Document)
I. Queries raised by M/s APH MEDICAL SYSTEMS		
Query No. 1	With the innovative Limb Occlusion Pressure (LOP) feature	The limb occlusion Pressure (LOP) feature which has been asked for is the latest and advanced patient safety feature where in personalized pressure for each patient needs to keep the tourniquet pressure to the minimum thereby reducing the risk of complications. We would like to procure the equipment as per Technical Specification floated through this tender.
Query No. 2	Ability to provide a specific Recommended Tourniquet Pressure (RTP) for each patient based on physiological characteristics.	
Query No. 3	Limb Occlusion Pressure (LOP) feature	
Query No. 4	Dedicated pressure line and pressure measurement line for cuff "one" and cuff "two" which should be designed to provide most accurate readings.	
Query No. 5	Should sense, calculate and report the cuff pressure necessary to achieve complete blood occlusion in the operative limb.	
Query No. 6	Should be US FDA, CE (European), UL or BIS approved	

The above minutes of Pre-Bid Clarifications is recommended for uploading for prospective Tenderer information on the AIIMS Bhopal e-Tendering Portal and websites. The Minutes of Pre-Bid Clarification is the form part of the referenced e-Tender document and changes indicated above are applicable and prevail for entire bid document.


(Dr. John A. Santoshi)
 Associate Professor
 Deptt. of Orthopedics

Dr. JOHN A. SANTOSHI
 Reg.No.MP11493 MSOrth,DNB orth,F Hand Surg