



No.: Stores/2016-17/LPC/.....

Dated: 07/12/2016

<b>Department:</b> Microbiology <b>Ref:</b> AIIMS/BPL/MB/16/909
--

Subject: Quotation for Supply of Medical Equipments.

Dear Sir,

Sealed quotation are invited from **local Vendors of (Bhopal)**, for supply of Medical Equipments with following Technical Specification attached Annexure "A" shown in enclosed bid so as to reach the **Stores Officer, Central Store, First Floor-Medical College Building AIIMS Bhopal by 14:00 Hrs of :14/12/2016**

**Terms & Conditions:-**

1. The quotation is to be submitted only in the format enclosed. Not acceptable in any other form.
2. The quotation must be submitted **in sealed envelope to the office of the Store officer (Central Stores Medical College) AIIMS Bhopal.**
3. Each page of the quotation must be printed on letter head of the Participants duly numbered.
4. Failure to comply with any of these conditions will lead to rejection of the quotation and Responsibility for the same will lie with the participants submitting the quotations.
5. **F.O.R.:** Rate quoted shall be fixed and on FOR destination Basis.
6. **Tax:** (i) Quotation must contain VAT/TAN/PAN Number otherwise the quotation will not be considered.  
(ii) Rate of applicable Tax should be shown separately.
7. **Delivery Period:** Supply is required immediately from ready stock but Maximum within a period of 21 Days.
8. **Payment:** Payment will be made within reasonable time from the date of acceptance of supply.
9. **Penalty:** If the supply is not received within the Scheduled delivery period, a Penalty @ 0.5% per week or part there of subject to a maximum of 5% shall be imposed which will be deducted from the Invoice.
10. **Right of rejection:** The Competent authority reserves all rights to reject any or all quotation without assigning any reason.
11. **Warranty:** **Warranty should Cleary stated on the quotation.**

## Technical Specifications Sheet

### “Annexure “A”

S. No.	Name of Item	Technical Specification
01.	<b>Analytical Balance, High Precision</b>	<ol style="list-style-type: none"> <li>1. Capacity: 220gms</li> <li>2. Readability: 0.1 mg</li> <li>3. Repeatability: <math>\pm 0.1</math>mg</li> <li>4. Linearity: <math>\pm 0.2</math>mg</li> <li>5. Circular removable pan of minimum 90mm diameter</li> <li>6. Backlit Digital Display LCD</li> <li>7. Glass Draft shield (3 sides) with doors which can be opened from sides and top</li> <li>8. Operating temperature 0- 45 ° C</li> <li>9. Calibration: Built in internal Calibration One standard certified weight for external calibration to be provided, along with calibration certificate from an NABL accredited Laboratory</li> <li>10. Should meet CE/FDA/BIS electrical compliances and safety compliances</li> <li>11. Should have overload protection</li> <li>12. Leveling feet with leveling indicator</li> <li>13. Operational manual to be provided</li> <li>14. Minimum three years warranty, 5 yrs CMC after the warranty period</li> <li>15. Power Supply: Power input to be 220-240VAC, 50Hz with plug compatible with local sockets</li> </ol> <p>01 Nos.</p>
02	<b>Vortex mixer</b>	<ol style="list-style-type: none"> <li>1. With brushless motor</li> <li>2. Suitable for Interchangeable mixing Heads</li> <li>3. Heavy metal base and rubber feet to prevent movement of the shaker during use</li> <li>4. Include one hand insert, micro tube insert &amp; cup attachment</li> <li>5. Maximum speed up to 3000 rpm with variable speed control knob</li> <li>6. Choice of continuous and touch mode working</li> <li>7. Operational manual to be provided</li> <li>8. Minimum 5 years warranty, 5 yrs CMC after the warranty period.</li> <li>9. Power: 220-240VAC, 50Hz with plug compatible with local sockets</li> </ol>
03	<b>Micro Centrifuge (Mini Spin)</b>	<ol style="list-style-type: none"> <li>1. Small foot print</li> <li>2. Should have independent ON/OFF switch for operation</li> <li>3. 6 place rotor for 1.5/2.0 ml tube</li> <li>4. 2 place rotor for 8 well strip tubes of 200 <math>\mu</math>l (0.2ml)</li> <li>5. Speed should be 6000 rpm (2000g)fixed</li> <li>6. Suitable base which can prevent movement of the equipment while working</li> <li>7. Operating temperature 0- 45 ° C</li> <li>8. Aerodynamic compact construction for vibration- and noise- free operation.</li> <li>9. Operational manual to be provided</li> <li>10. Minimum 5 years warranty, 5 yrs CMC after the warranty period</li> <li>11. Power: 220-240VAC, 50Hz with plug compatible with local sockets</li> </ol>

## Financial Bid (On Letter Head)

Ref No.:

Name of Bidder:

Date:...../12/2016

Full Address of the Company:

E Mail Id:

Tin No:

Sr.No	Description	Technical Specifications	Req. Qty.	Brand/Make	Model No:	Unit Price in Rs.	Discount if any.	Net Unit Cost after Discount	Total Cost in Rs.
1	2	3	4	5	6	7	8	9 = 7-8	10 = 9 x 4
1	Analytical Balance, High Precision	As per Annexure "A"	1						
2	Vortex mixer	As per Annexure "A"	1						
3	Micro Centrifuge (Mini Spin)	As per Annexure "A"	1						
<b>Sub Total (A)</b>									
<b>VAT / CST Extra @ Rs.</b>									
<b>Total (inclusive of all ) taxes in Rs .</b>									

**Signature with Stamp**

**Place:**

**Date:**