

All India Institute of Medical Sciences
Saket Nagar Bhopal

Pre-Bid Clarification and Corrigendum

AIIMS/BPL/ANAT/ 1047

Dated 31 May.2016

With reference to Tender No. 229 AIIMS BHOPAL/HC/RC-Anatomy/Anthropology,Osteology & Cytogenic Lab. Items/2015-16/01 Dated :21.04.2016 for Notice ,Inviting Tender for Supply, Installation and support of Anatomy/Anthropology,Osteology & Cytogenic Lab. Items

Subsequent to Pre-Bid queries sought for clarification in Pre-Bid meeting or queries received through e-mail/hard copy ,the following clarification to Pre-Bid queries are circulated for the wider information to the prospective tenders and requested them to download Pre-Bid clarification and submit the same along with their "Technical Bid" envelope .


Queries raised by

1.M/S DSS Imagetech Pvt. Ltd , New Delhi

S.No	Tender Specification	Query	Clarification
1	Digital CCD camera with high sensitivity and low noise. <input type="checkbox"/> Chip size should be 2/3" <input type="checkbox"/> High resolution of 1360x1024 pixels with pixel size of 6.45 x 6.45um <input type="checkbox"/> Digitization depth – 12 bit <input type="checkbox"/> Frame rate of 15 frames per second (fps) in full resolution <input type="checkbox"/> C-mount adapter 0.63x <input type="checkbox"/> FireWire Interface	Fire Wire Interface/Gigabit Interface	Fire Wire Interface/Gigabit Interface
2	Warranty 7 yrs for instruments & 10 yrs for software	Warranty 2 years for instruments & software.AMC/CMC ager warranty period for five years should be quoted separately	Warranty 3 years for instruments & software. AMC/CMC for 5 yrs after expiry of 3 yrs warranty

Corrigendum

S.No.	Specification	Recommendation
1	Group No. 3	All standard accessories applicable for smooth full functioning of equipments
2	Frequency of visits during Warranty/CMC	One visit every three months (4 visits in a year) for periodic/preventive maintenance and any time for attending repairs/break down calls


31.05.16

Dr. Bertha AD Rathinam
Professor & Head
Department of Anatomy
AIIMS Bhopal