



Prebid clarification for tender no 152 : AIIMS BHOPAL/HC/RC-PULMONARY/NON INVASIVE VENTILATOR, COUGH ASSIST DEVICE, PORTABLE VENTILATOR/2014-15/01 Dated: 11.10.2014

	Existing	Corrigendum
Non invasive ventilator		
1.	Should have built in internal battery for minimum 5 hours of back up time.	Should have built in internal battery for minimum 2 hours of back up time
2.	It should have an LCD monitor with display atleast 4 cm for real-time monitoring to track patient response	It should have an LCD monitor with display atleast 4 inch for real-time monitoring to track patient response
3.	Backup Respiratory rate 0 to minimum 30 BPM	Backup Respiratory rate 0 to minimum 40 BPM
Portable ventilator		
4.	Point no. 13 Should be USFDA and European CE approved	Should be USFDA or European CE approved


 Dr. Abhishek Goyal
 Asst Professor
 AIIMS Bhopal
 Department of Pulmonary Medicine
 Reg. No. DMC / R-02490
 All India Institute of Medical Sciences
 Saket Nagar, Bhopal-462024


 Dr. Aikeshi Khurana
 Asst Professor
 AIIMS Bhopal