

Prebid Corrigendum Tender No.:181

AIIMS BHOPAL/HC/RC/FIBEROPTIC BRONCHOSCOPE, VIDEO BRONCHOSCOPE & ARGON PLASMA COAGULATION SYSTEM /2014-15/01

For all items 1-3: Tender opening will be done on 3 Feb 2015 .		
	Existing	Corrigendum
<b>Fiberoptic Bronchoscope</b>		
1.	It should be USFDA or European CE approved.	Must have USFDA or CE certification
<b>APC</b>		
2.	Modes of operation: Foot switch	Modes of operation: Foot switch (Two in Nos.)
3.	Electro Surgical unit should have three modes for APC ,Pulsed and Forced	Electro Surgical unit should have three modes for APC : Precise, Pulsed and Forced
4.	Coagulation mode should have option of variable coagulation like soft coagulation, forced coagulation, spray coagulation	Coagulation mode should have option of variable coagulation like soft coagulation and forced coagulation
5.	It should be upgradable to APC ( with two modes)	It should be upgradable to APC ( with three modes)
6.	Should have two different modes of APC.	Should have three different modes of APC.
7.	Different catheters: Length 1.5 meter to 3.0 meters 1.5-2.5 meters (Straight beam) - 05 Nos. 1.5-2.5 meters (side conical) 05 Nos. 3.0 meters (Straight) 02 Nos.	Different catheters: 1. Rigid applicators: Length 240 to 300 – 3 Nos 2. 1.5-2.5 meters (Straight beam) - 05 Nos. 3. 1.5-2.5 meters (side conical)- 05 Nos. 4. 3.0 meters (Straight)- 02 Nos.
8.	Irrigation pump for endoscopic procedures	This point should be omitted.

  
Dr Abhishek Goyal

Asst Professor

AIIMS Bhopal