

Minutes of Pre-Bid Clarifications

Amendment No. 1

Page ① of ②

Dated: 26.06.2018

The Pre-Bid Meeting was held on 04.06.2018 at 15:00 Hrs. in the MS Office Annex., AIIMS Bhopal Hospital with reference to AIIMS Bhopal e-Tender Enquiry No. AIIMSBPLHOSP181901925052018 for E-Tender for Rate Contract for Supply, Installation, Testing and Commissioning of DVT Pump for Department of Orthopaedics at AIIMS Bhopal Hospital.

Following Interested Tenderers representatives were participated in Pre-Bid Meeting and submitted their query through official e-mail/ in-person etedner@aiimsbhopal.edu.in as follows:-

Sr. No.	Prospective Bidder Company Name	Pre-Bid Query Letters enclosed at
1.	M/s Indian Surgical Equipment Co. Pvt. Ltd.	Flag "A-1"

The Pre-Bid Queries Clarification applicable for this e-Tender is as under:-

S. No.	Prospective Bidder	Query No.	Tender Clause	Representation Received from the Bidder	Tender terms and conditions amended as under (whereas applicable in entire Tender Document)
1	M/s Indian Surgical Equipment Co. Pvt. Ltd.	1	Tender Item Sr.No. 1 DVT Pump Technical Specification/ Description of the requested Equipment Annexure I at Pg. No. 20:-Increases blood flow – Velocity by over 220% on resting mode.	Tender Item Sr. No. 1 DVT Pump Technical Specification/ Description of the requested Equipment Annexure I at Pg. No. 20 :- These are never a part of Standard DVT Pump specification and seems specific to one company. This point may be removed	The specifications of the tender are modified to read as under: DVT Pump 1. Should provide graduated, sequential compression and rapid impulse inflation to calf, foot & thigh. 2. Should have Pulse frequency 1 per minute range. 3. Should have choice of three cuffs of universal size: Calf, thigh, foot 5. Should deliver constant pre-set pressure ranges Distal 40-130 mmHg and Proximal
		2	Tender Item Sr.No. 1 DVT Pump Technical Specification/ Description of the requested Equipment Annexure I at Pg. No. 20 :- Provides graduated, sequential compression and rapid impulse inflation to calf, foot & thigh.	Tender Item Sr. No. 1 DVT Pump Technical Specification/ Description of the requested Equipment Annexure I at Pg. No. 20 :- Choice of three cuffs of universal size: Calf, thigh, foot Please clarify whether you require disposable cuffs or Reusable.	

(Signature)
27/7/18

(Signature)

(Signature)
27/7/18

S. No.	Prospective Bidder	Query No.	Tender Clause	Representation Received from the Bidder	Tender terms and conditions amended as under (whereas applicable in entire Tender Document)
		3	Tender Item Sr. No. 1DVT Pump Technical Specification/ Description of the requested Equipment Annexure I at Pg. No. 20 :- Should deliver constant pre-set pressure ranges – Distal 52 pulse mince 10 % mmHg	Tender Item Sr. No. 1DVT Pump Technical Specification/ Description of the requested Equipment Annexure I at Pg. No. 20 :- Should deliver constant pre-set pressure ranges-Distal 52 pulse mince 10% mm HG. This is specific to one company and may be amended as: Proximal 30-50 mm Hg	30-50 mmHg. 6. Alarm Should be present 7. Should have visual display screen for pressures and time present 8. Should be portable, can be mounted on the bed. 9. Should provide battery backup at least 4 hours. 10. Should have US-FDA/ European CE with 4 digit notified body number approved product. 11. The cuffs should be reusable. 12. Vendor to provide rates for both disposable as well as reusable cuffs separately.
		4	Tender Item Sr. No. 1DVT Pump Technical Specification/ Description of the requested Equipment Annexure I at Pg. No. 20 :- Pro 45 pulse mince 10 % mmHg	Tender Item Sr. No. 1DVT Pump Technical Specification/ Description of the requested Equipment Annexure I at Pg. No. 20 :- Pro 45 pulse mince 10% mmHg This is specific to one company and may be amended as: Proximal 30-50 mmHg	

The rest of all terms and conditions floated will remain same as depicted in e-tender Document.
The above Minutes of Pre-Bid Clarifications and Amendment No. 1 will be part of the e-Tender Document and changes/addition/deletion above in Pre-Bid Clarifications and Amendment No. 1 will be applicable.

Vivek
27/7/18

Dr. M. Nagar
27/7/18

(Dr John A. Santoshi)
Addl. Professor
Department of Orthopaedics
All India Institute of Medical Sciences
Saket Nagar, Bhopal (M.P.)-462020