

Pre-Bid Clarification			
S.NO.	Item Name.	Tender Specification	Queries raised By KARL STORZ
1	Pediatric Fiber Optic bronchoscope Set	Equipment should be US FDA/European CE/BIS/ISO approved and the company should be ISO 9001 certified	ISO 9001 Certification should be removed as it is for processing certification and not for quality certification.
2	Pediatric Fiber Optic bronchoscope Set	Monitor: 15 inches HD monitor of medical grade. It should be mountable on trolley	Monitor size should be 17-19 inches
3	Pediatric Fiber Optic bronchoscope Set	Monitor: 15 inches HD monitor of medical grade. It should be mountable on trolley	Monitor: 17-19 inches HD monitor of medical grade. It should be mountable on any arm.
4	Pediatric Fiber Optic bronchoscope Set	BOQ: Bill of Quantity	Pediatric FiberOptic - 02 Imaging System Monitor & Trolley - 01
5	Pediatric Fiber Optic bronchoscope Set	Trolley	Trolley should be of good quality & make.
6	Pediatric Fiber Optic bronchoscope Set	Specification need to be added	System should be able to record video with inbuilt/external storage of atleast 2 GB.
7	Pediatric Fiber Optic bronchoscope Set	Depth of focus should be 3-50mm	Depth of focus should be as per standards.
8	Video Laryngoscope Set	Cost and Quantity	BOQ: Bill of Quantity:- 02 Video Laryngoscope with all the Blade size.
9	Video Laryngoscope Set	Accessories like protection cap, tray for cleaning and sterilization of blades (at least two blades at a time) should be provided.	01 Imaging System. Accessories like protection cap, tray for cleaning and sterilization of blades (at least two blades at a time) should be of same Make.
10	Video Laryngoscope Set	One Imported Trolley to hang Scope as well monitor manufactured by same company should be provided.	One Trolley to hang FiberOptic bronchoScope as well monitor of any good Make should be provided.
S.NO.	Item Name.	Tender Specification	Queries raised By Olympus
1	Pediatric Fiber Optic Laryngoscope Set	Pediatric Fiber Optic bronchoscope Set & Video Laryngoscope Set	We request you to kindly split EMD for both item no.1 & Item no. 2
2	Pediatric Fiber Optic bronchoscope Set	Distal and diameter should be 2.7 mm - 2.8 mm	We request you to kindly amend it as..... Distal and diameter should be 2.7 mm - 3.1 mm

Amrta Tripathi

Jess
31.03.2020

Amrta