



अल भारतीय आयुर्विज्ञान संस्थान (एमएस) भोपाल  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES BHOPAL  
E-Tender No. AIIMS/BPL/HOSP/2020-21/121

(Flag - A)

	Tender Document	Vendor Quarries :- Vendor 1(Baxter India Pvt ltd. )	Bid Clarification
1	Annexure-V	<p>Specifications of Item</p> <p>4. Low flux dialyzer ,SA 1.8m2,Polysulphone membrane, Polycarbonate housing, Sterlization steam,UfCo-18,Urea cl-252 at Qb 300</p> <p>6. Low flux dialyzer ,SA 1.3m2,Polysulphone membrane, Polycarbonate housing, Sterlization steam,UfCo-13,Urea cl-243,priming vol 78 at Qb 300</p> <p>Polyarylethersulphone membrane have very high clearance and endotoxin retension(KoA more than 825) hence allow polymix membrane dialyzer along with other membranes.</p> <p>And to consider Surface area from 1.7-1.8 m2 in item no-4and 1.3 to 1.4 m2 SA in item no 6.</p>	<p>4. Low flux dialyzer ,SA 1.7-1.8m2,Polysulphone/Polymix membrane, Polycarbonate housing, Sterlization steam,UfCo-18,Urea cl-252 at Qb 300 ml/min</p> <p>6. Low flux dialyzer ,SA 1.3-1.4m2,Polysulphone/polymix membrane, Polycarbonate housing, Sterlization steam,UfCo-13,Urea cl-243 or above, at Qb 300ml/min</p>

*Handwritten signature and date:*  
29/5/2020  
A. M. Allan