



Date:- 31/08/2021

Pre Bid Clarification

Sub:E-TenderNo.-AIIMS/BPL/MC/MICROBIOLOGY/21-22/160,Dated:02/07/2021

E-Tender documents for Supply, Installation, Testing and Commissioning of Automated Anaerobic Jar Gas Filling System for Microbiology Department, AIIMS Bhopal.

S. No.	Reference	Stipulated Tender Condition	Query Raised by M/s Yunymidex, Delhi	Amendment in technical Specification of equipment in tender document
1	As per Tender Item Specification Serial No. 5	System should work with standard transparent or stainless steel anaerobic jars of any make	System should work with either Polycarbonate or Stainless Steel Jars	No Change
2	As per Tender Item Specification Serial No. 10	System to be offered with 10 plate polycarbonate & 48 plate stainless steel jars	In Specification no. 10, It is mentioned that the equipment should work with Both Polycarbonate and Stainless Steel Jars. This makes the specification less competitive as equipment dealers who can only supply Polycarbonate Jars along with their equipment will not able to quote for the Tender or would either get rejected so the Specification No 10 – should be reframed as Either Polycarbonate or Stainless Steel Jars. Stainless steel jars are less user friendly because they are not transparent, the user cannot assess the bacterial growth inside the jars without opening jars, whereas a polycarbonate or methacrylate jar the growth can be seen from outside without opening the jar.	System to be offered with Polycarbonate OR Stainless Steel Jars.

3	As per Tender Item Specification Serial No. 12	Should be ISO 9001:2015 quality management system for microbiological applications, UKAS accreditation & should conform to EN IEC 61010, EC Directive 73/23/EEC, 89/336/EEC and 89/392/EEC	<p><u>Specification No.12</u> – regarding equipment certificates. These directives are also restrictive.</p> <p><u>THE UKAS Directive</u> should be Mentioned as an option and USFDA should also be given as option.</p> <p><u>THE ISO 9001:2015</u> is a Universal Directive so it could be included as a must for Quoting the Tender.</p> <p><u>The Three EC Directives – 73/33/EEC, 89/336/EEC and 89/392/EEC</u> – Should be kept as an option as these are applicable for European Member States Only.</p> <p><u>EN IEC 61010</u> – Is as International Directive and should be kept as Mandatory.</p>	Should be ISO 9001:2015/USFDA/CE/UKAS certified; the supplied system confirm to EN IEC61010 safety standard.
---	--	--	---	--

Sd/-
Prof. I/c Central Stores
AIIMS Bhopal