


Date: 21.09.2020

Subject: pre bid clarifications and suggestion for e-tender no- E-Tender Id:-AIIMS/BPL/HOSP/
NIBP/ PHARMACOLOGY/20-21/127 Dated: 03/09/2020 (Non Invasive Blood Pressure (NIBP)
Monitoring System for Rodents).

Replies to the pre-bid queries raised by M/s Technical Systems, Bhopal.

Sl. no.	Tender Specification (Chapter-IV)	Request Changes if any with reason	Decision of TSC
1.	The system should be compact and can be used for rats	A single system can be used for both rats and mice.	May be read as "NIBP monitoring system can be used for both rats and mice".
3.	Should have built-in Warmer for maintaining temperature.	Warming pads are externally used for maintaining temperature and rectal probes for monitoring the temperature.	No Change
5.	Should be able to display real-time parameters- Systolic, Diastolic, Mean, Heart rate, and Temperature.	More parameters can be measured. Analysis software required for further required for further analysis.	No Change
7.	Should have USB port for direct data transfer.	Data Should be transferred to software through USB for recording and analysis. It should have facility for online and offline analysis.	No Change.

The decision of the TSC is appended herewith.


I/c, SPSO, Central Stores
AIIMS Bhopal


31-08-20


NP-11


The technical specification committee recommended that requested change by Technical system (7d/15) the requested change S.No. 1 that "NIBP monitoring system can be used for both rats & mice" to be accepted.


Other requested changes S.No. 3, S.No. 5 & S.No. 7 were not accepted by the technical specification committee

AIMS BPL / PHARM / 20 / 25 / No. 688 / Dt. 15.9.20
2020/15/20


(Dr. Rajy Shashankar)


(Dr. Latinder Thaj)


(Dr. Babukrishnan S)


(Dr. Ajay Shukla)

15.9.2020