



अखिल भारतीय आयुर्विज्ञान संस्थान
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
सकेत नगर भोपाल (मध्य प्रदेश) - 462020
Saket Nagar, Bhopal (M.P.) - 462020
website: www.aiimsbhopal.edu.in

AIIMS, BHOPAL
CENTRAL STORES DEPARTMENT
Dispatch No. 1919
Date: 28/1/23

Date:- 28/01/2023

Pre Bid Clarification & Bid Submission Date Extension

Sub:E-TenderNo.- AIIMS/ BPL/ HOSP/ CIFRAPID AUTOCLAVE/2021-22/186,
Dated:22/12/ 2022

E-Tender documents for Supply, Installation, Testing and Commissioning of Rapid Autoclave to Central Instrumentation Facility department at AIIMS Bhopal.

Pre Bid Clarification/Amendments are being uploaded for information of the prospective Bidders. All prospective Bidders are requested to prepare and submit their Bids, accordingly.

These Clarifications/Amendments will form part of above mentioned E-Tender ID. Bids not found submitted as per these clarifications/approved amendment, shall be liable for rejection.

Consequent upon these clarifications, the date of the submission of bids is also extended as particularized below:-

Events

Revised Dates

E- Tender Submission End Date	:	14/02/2023	17:00 Hrs
E- Tender Technical Bid Opening Date	:	15/02/2023	15:00 Hrs

Other terms and conditions will remain the same.

Venue for offline tender documents submission: Tender box kept in the Central Stores department, 1st Floor, Service Building-II, AIIMS Bhopal-462020, (MP.)

I/c Senior Procurement cum Store Officer
Central Stores
AIIMS Bhopal

31/01/23

141

• Technical Specifications amendments

Name of the instrument: Rapid autoclave

Specs.Sr. No.	Tender Specification	Our Request for Amendment	User dept. Remarks
1.	Portable High-Speed microprocessor-controlled rapid autoclave suitable for sterilization of unwrapped instruments, wrapped instruments, Packs and special cycle for liquids.	Portable High-Speed microprocessor-controlled rapid autoclave suitable for sterilization of unwrapped instruments, Wrapped instruments.	Change in the tender specification not accepted as equipment will also be used for microbiological usage therefore without cycles for liquids not fulfil the objective of procurement.
2.	Have 2 or more programmable cycles to allow custom creation of different cycle parameters for special sterilization needs. The programmable cycle should allow, change of time, temperature, Dry time and Vent.	Have one programmable cycle to allow custom creation of different cycle parameters for special sterilization needs. The programmable cycle should allow change of time, temperature, Dry time and Vent.	Not Accepted. As per need of multiple users it should have 2 or more programmable cycles.
3.	Pre-programmed cycles should have selection of 2 temperatures (132°C and 121°C) and pressures (186 kPa and 104 kPa).	Pre-programmed cycles should have selection of 2 temperatures (134°C and 121°C) and pressures (186 kPa and 104 kPa).	Accepted with pre-programmed cycles should have selection of 2 temperatures (132°C-134°C and 121°C) and pressures (186 kPa and 104 kPa).
6.	Should meet the requirements of ASME Boiler and Pressure Vessel Code.	Should meet the requirements of ASME Boiler and Pressure Vessel Code/ Compliance of Chamber for Chamber Pressure Equipment Directives (PED) AND EN13445 norms	Accepted

Handwritten signature and date: 24/1/23

Handwritten signature and date: 24/1/23

Handwritten signature and date: 24/01/2023

Handwritten signature and date: 24/01/2023

Handwritten signature and date: 24/01/23

Handwritten signature and date: 25/1/2023

7.	Should be ISO 9001 certified, FDA certified, CSA certified	Should be ISO 9001 certified/FDA certified/ CSA certified	Accepted. It Should be ISO 9001 certified/ FDA certified/ CSA certified.
8.	Chamber size should be 27 cmdiameter or more and chamberdepth should be 45 cm or more.	Chamber size should be 25 cmdiameter or more and chamberdepth should be 35 cm withvolume 17 liters or more.	Not Accepted. Reducing of the chamber diameter and depth will reduce the volume capacity of the autoclave. Hence, the requested amendment not accepted.
19.	Instrument shall be supplied withall the accessories required to function.	Instrument shall be supplied with all the accessories required to function including: Water purification system of suitable capacity, Printer Paper- 10 rolls and Stainless-Steel Table for placement of Autoclave.	Accepted. The requested accessories should be supplied with the equipment of well-established Indian make as per the AIIMS Bhopal terms and conditions.
Clause No	Terms as per Tender	Our Request for amendment	Remarks
Chapter-1, Clause3	Price Preference Policy and Exemption for submission of various eligibility Criteria documents to the BIDDER Registered under Make in India Initiative.	May please be removed.	Not accepted. Since the product name is not mentioned in MII exemption list in which relaxation on Global Tender Enquiry (GTE) under Rule 161(iv) of General Financial Rules (GFRs) 2017 is given. Therefore, request to remove the clause is not accepted.

Chakraborty
24/01/23

7.7
24-01-22

Meghate
24/01/23

Sharma
24/1/23

Rupinder
25/1/2023