


Response to Pre bid Queries (Pre bid date: 10/12/2019)					
E Tender ID : AIIMS/BPL/HOSP/2019-20/103					
Rate contract for supply, installation, commissioning and testing of Patient Warming System, Department of Paediatric Surgery, AIIMS Bhopal					
S. No	Tender Page and Para	Tender Specifications	Name of the firm	Representation from the firm	TSC Recommendation
1	Page 37, Para 3	Should be based on semiconductor polymer foil for precise warming of entire patient body during & after surgery.	Indian Surgical Equipment Co. Pvt. Ltd	Should be based on semiconductor/ carbon fiber polymer foil for precise warming of entire patient body during & after surgery.	Accepted
2	Page 37, Para 5	Control unit should be capable of warming minimum four segments at a time.	Indian Surgical Equipment Co. Pvt. Ltd	Control unit should be capable of warming minimum 2-4 segments at a time.	Accepted
3	Page 37, Para 7	Control unit should have touch screen display to select & display temperature of all four segments at a time.	Indian Surgical Equipment Co. Pvt. Ltd	Control unit should have touch screen display to select & display temperature of all segments at a time.	Accepted
4	Page 37, Para 9	Should offer precise digital temperature control with selectable temperature range of 30 to 40° C in steps of 0.1° C	Indian Surgical Equipment Co. Pvt. Ltd	Should offer precise digital temperature control with selectable temperature range of 33 to 39° C in steps of 0.1° C	No Change required
5	Page 37, Para 10	Arm cum shoulder segments should be divided in two sections capable of being switched ON or OFF independently depending upon the nature of surgery and condition of patient	Indian Surgical Equipment Co. Pvt. Ltd	Arm cum shoulder segment should be in single section.	No Change required
6	Page 37, Para 12	Should also have on screen graphical display of patient body temperature for the entire duration of surgery.	Indian Surgical Equipment Co. Pvt. Ltd	Should also have on screen graphical/ digital display of patient body temperature for the entire duration of surgery.	Accepted


Dr. REYAZ AHMAD
 M.S., M.Ch (Paediatric Surgery)
 Associate Professor
 Department of Paediatric Surgery
 AIIMS, Bhopal.


HOD
 Department of General Surgery
 AIIMS, Bhopal-462020 (M.P.)


Dr. MAHENDRA KUMAR JANGID
 M.S., M.Ch (Plastic Surgery)
 Associate Professor
 Deptt. of Burns & Plastic Surgery
 AIIMS, Bhopal (M.P.)-462020