



अखिल भारतीय आयुर्विज्ञान संस्थान
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
सकेत नगर भोपाल (मध्य प्रदेश) - 462020
Saket Nagar, Bhopal (M.P.) - 462020
website: www.aiimsbhopal.edu.in



Date:- 24.01.2023

Pre Bid Clarification & Bid Submission Date Extension

Sub:E-TenderNo.-AIIMS/BPL/MC/MICROBIOLOGY/21-22/181, Dated:13/12/2022

E-Tender documents for Supply, Installation, Testing and Commissioning of Automated Immunoassay System for the department of Microbiology Department, AIIMS Bhopal.

Pre Bid Clarification/Amendments are being uploaded for information of the prospective Bidders. All prospective Bidders are requested to prepare and submit their Bids, accordingly.

These Clarifications/Amendments will form part of above mentioned E-Tender ID. Bids not found submitted as per these clarifications/approved amendment, shall be liable for rejection.

Consequent upon these clarifications, the date of the submission of bids is also extended as particularized below:-

<u>Events</u>	<u>Revised Dates</u>
E- Tender Submission End Date	: 08/02/2023 17:00 Hrs
E- Tender Technical Bid Opening Date	: 09/02/2023 15:00 Hrs

Other terms and conditions will remain the same.

Venue for offline tender documents submission: Tender box kept in the Central Stores department, 1st Floor, Service Building-II, AIIMS Bhopal-462020, (MP.)

I/c Senior Procurement cum Store Officer
Central Stores
AIIMS Bhopal

Sr. Procurement Cum Store Officer
Central Stores Department
AIIMS, Saket Nagar, Bhopal (M.P.)

Name of Bidder participated in the pre-bid meeting	Suggestions/Queries raised by the Bidder	Decision of the departmental TSC Members
(1) Ms/Channel Ten	i. EIA Technology	TSC decided to adhere to the original technical specifications vetted and submitted by the committee.
	ii. Calibrators & Controls to be provided with the kit	
	iii. Calibration & QC to be lot-wise (So no hassle of maintaining LJ plot)	
	iv. Pack size - 24 tests (minimum)	
	v. Infectious markers panel of over 40 parameter including viral, bacterial, parasital marker.	
	vi. 30 Sample position	
	vii. Monotest cartridge based system	
(2) M/s Roche Diagnostics India Pvt. Ltd	i. Automated Immunoassay system estimated cost is too low, request you to increase value approx 25 to 45 lac + tax	The decision to raise the estimated cost of equipment will be as per the rules of the institute.
	ii. Clostridium difficile toxin detection assay are not available on our platform	Clostridium difficile toxin detection assay is a routine diagnostic test and hence will be required.
Ortho Clinical Diagnostic	i. Budget of the tender needs to be increase as it has clause of 5 years warranty, minimum of 50 lacs as you have asked for FDA approved.	The decision to raise the estimated cost of equipment will be as per the rules of the institute.
	ii. Request you to kindly make optional test for avidity test & Clostridium difficile toxin test	These tests can't be made optional.
	iii. Throughput of instrument needs to be increased minimum of 50 test/Hour	TSC decided to adhere to the original technical specifications vetted and submitted by the committee
	iv. Onboard sample needs to be increased from 10 to minimum of 50 as hospital is asking for random analyzer.	
	v. UPS along with instrument as per specifications, and as discussed hospital have central UPS line so company UPS will be kept idle, so request you to please look into it (we supply third party for it)	
Clarification for the clause 'The free of cost consumables (if any required) for 3 months' as mentioned on folio no. 133 (point no.5)		TSC suggested making this clause optional for the participating vendors in the tender bid.

Shashank
14/1/23
Dr. Shashank Purwar

Asok
14.01.23
Dr. Asok Kumar

Ratinder
14/1/23
Dr. Ratinder Jhaj

Shwati
14/01/2023
Dr. Shashwati Nema