



अखिल भारतीय आयुर्विज्ञान संस्थान  
**ALL INDIA INSTITUTE OF MEDICAL SCIENCES**  
साकेत नगर भोपाल (मध्यप्रदेश) – 462020  
Saket Nagar, Bhopal (M.P.) – 462020  
Website : [www.aiimsbhopal.edu.in](http://www.aiimsbhopal.edu.in)

**Ref. No.: Stores/PAC-Rule -166(i) /20-21/152**

**Dated:06/06/2020**

**Subject:** Proposal for the procurement of "**Eye Bank Specular Microscope model Cell Check D**" for the Ophthalmology Department, Hospital Complex, AIIMS Bhopal, Saket Nagar Bhopal -462020

The request received from Ophthalmology Department, AIIMS Bhopal for the purchase of captioned Item **from M/s BioMedix Optotechnik & Devices P Ltd., on Proprietary basis.**

The Notice is being uploaded for general information of Aspirant Manufacturer/Dealer/Distributor to submit their objections/proposal, if any, on proprietorship of these items.

In case, the product of any Manufacturer/Authorized distributor/dealer conforms to the enclosed specifications, they may submit their proposal for the supply same item along with the following:-

- (i) Product brochure;
- (ii) Point-by-point compliance of the enclosed specifications, along with all relevant documentary evidence.

**The objection/ proposal** should be sent in sealed cover to, The **Sr. Procurement Officer cum Stores Officer, Hospital Complex Building, Third Floor, AIIMS, Saket Nagar, Bhopal (462020)** so as to reach on or **before ..../06/2020 upto12:00 Hrs.**, failing which it will be presumed that no any other vendor is interested to offer comments/protest and case will be decided accordingly on its merit.

The reference number: **Stores/PAC-Rule -166(i)/20-21/152, Date: 06/06/2020, due on: 27/06/2020** should be super scribed on sealed envelope.

**Yours faithfully**  
**I/c Central stores**

- Encl: 1. Authorization Certificate**  
**2. PAC Certificate**  
**3. Schedule of requirement**

**Authorization Certificate**

**KONAN™**  
MEDICAL

**KONAN MEDICAL, INC.**

10-29, Miyanishicho, Nishinomiya, Hyogo 662-0976, JAPAN  
TEL +81-798-36-3456 FAX +81-798-26-1028

**Authorization Letter**


To whom it may concern,

We, Konan Medical, Inc. 10-29 Miyanishicho, Nishinomiya, Hyogo 662-0976, Japan, hereby certify that M/s. Biomedix Optotechnik & Devices Pvt. Ltd., No. 258, Amarjyothi Layout, Domlur, Bangalore – 560 071, India, is our authorized distributor for the complete product range in India.

M/s. Biomedix Optotechnik & Devices Pvt. Ltd. is authorized by us to sell, bid, negotiate, distribute, enter into sales/service contracts and to be responsible for after sales service on behalf of us in India.

This Authorization Letter shall be valid until 31st March 2021.

Yours faithfully,

  
\_\_\_\_\_  
IKEGAMI, Tetsuji / President & CEO  
Konan Medical, Inc.

Date: 1st April, 2020

株式会社コーナン・メディカル

〒662-0976 兵庫県西宮市宮西町10-29 TEL 0798-36-3456(代表) FAX 0798-26-1028  
〒169-0075 東京都新宿区高田馬場4-40-17 TEL 03-5337-2025(代表) FAX 03-5337-2026

**PAC Certificate (Proprietary certificate)**

**KONAN™  
MEDICAL**



**KONAN MEDICAL, INC.**

10-29, Miyashicho, Nishinomiya, Hyogo 662-0976, JAPAN  
TEL +81-798-36-3456 FAX +81-798-26-1028

Date: 31 July, 2019

**PROPRIETARY CERTIFICATE**

**TO WHOM SO EVER IT MAY CONCERN COUNTRY OF INDIA**

This is to certify that we Konan Medical Incorporated, 10-29, Miyashicho, Nishinomiya, Hyogo 662-0976, Japan are the sole manufacturer of the Eye Bank Specular Microscope Model Cell Check D and BioMedix Optotechnik & Devices Pvt Ltd, 258 Amarjyothi Layout, Domlur, Bangalore, Karnataka – 560071 is exclusive authorized sales and service representative in the country of India. The proprietary features are as mentioned below.

The equipment is incorporated with center method, multi-area analysis and thermometer.

No other company other than Konan Medical Incorporated, Japan manufactures an Eye Bank Specular Microscope in the world with the above-mentioned feature.

Eiji Takaso

Sales Manager

Konan Medical, Inc.

株式会社コーナン・メディカル

〒662-0976 兵庫県西宮市宮西町10-29 TEL 0798-36-3456(代表) FAX 0798-26-1028  
〒160-0075 東京都新宿区高田馬場4-10-17 TEL 03-5337-2025(代表) FAX 03-5337-2026

**Schedule of requirement and brief specifications**

Sr.No	Description	Req qty.
1	2	3
1	<b>Eye Bank Specular Microscope model Cell Check D</b>	01 No

**Specifications as below:-**

The equipment should be reliable, advanced international standard machine with a long standing reputation

1. Eye bank keratoanalyzer
2. Ultra-Wide viewing field-0.48 x 0.6mm (analysis area-0.2 x 0.28 mm)
3. Dual camera system for finder and cornea-dual screen provided by two high resolution cameras.
4. Built-in thermometer-Display of the media temperature.
- 5 Built in pachymeter. Compact design dimensions: 200-220(W) x 225- 230 (H), X200-2200mm, weight:6.1 kg
6. Viewing field-Software allowing view of bro viewing field 0.48\*0.60 mm illumination
1. Display LCD- temperature and pachymetry chamber. Electrical AC 100-240V 50/60 HZ. 50/60 HZ, SOVA vial dimensions upto 35 mm diameter size 200 (W) X 255 (H)x220[Dmm, thermometer 0-45 degree Celsius, interface USB 2.0 stage ranges.
8. Cameras dual CMOS finder
9. Cornea operating conditions- ambient temp:10-40 degree Celsius, relative humidity 30%-85% atmospheric pressure: 70 Kpa to 105 Kpa
10. Motorised table-Heavy duty motorized table with slots for keeping slides and standard accessories.
11. The instrument should be with USFDA/CE or equivalent International certification
- 12 The instrument should be with online UPS or CVT as per standard recommendation and with 5 year replaceable warranty and dust cover

0

**\*\* Note: Please don't quote the prices of the above schedule requirement**