



अखिल भारतीय आयुर्विज्ञान संस्थान, भोपाल
ALL INDIA INSTITUTE OF MEDICAL SCIENCES (AIIMS) BHOPAL

Saket Nagar, Bhopal-462020 (India)
Website :www.aiimsbhopal.edu.in

Pre- Bid Clarifications

E-Tender No. 055 AIIMS/BPL/STORES/CIF/IN-VIVO/2019-20 Dated: 10.12.2019
(In-Vivo Animal Imaging System to the Central Instrumentation Facility)

1. M/s Icon Analytical Equipment Pvt. Ltd.

Reference	Tender Specifications	Request for Amendment	Remarks	AIIMS Bhopal Response
Point-1	Should be an automatic system to image small animals (mice, rats etc)	The number of mice/rats should be mentioned	This would help decide the suitable models to be quoted	Suggested amendment not accepted
Point-5	System should have min of 19 excitation and 7 emission filters. Both excitation and emission filter wheels should be motorized and software controlled	System should have suitable number of emission and excitation filters to fulfil imaging needs for the range mentioned in point no. 2 of tender.	Specific number of filters mention specific make.	Suggested amendment not accepted
Point-12	System should be compatible with animal ID chip system with individual identification number and captured IDs can be integrated into the system software.	This point should be removed from the tender	This is a proprietary technology used by a specific vendor which will not allow others to participate.	The specification is amended as System should have animal ID system for providing individual identification number and IDs can be integrated into the system software.
Point-18	Image Software: 1 acquisition copy and 4 analysis copies of Image software version 4.0 and higher should be provided with the system.	Image Software: 1 acquisition copy and 4 analysis copies of Image software of latest version should be provided with the system.	Instead of mentioning version 4.0 and higher, the latest version of the software should be mentioned. This will give scope to provide the latest software.	The specification is amended as 1 acquisition copy and 4 analysis copies of Image software with latest version should be provided with the system.

BB
01/01/2020

S. Kashyap
01/01/2020

RD
01.01.2020

ASR
1.1.20

RD
1/1/2020

AA
1/1/20



अखिल भारतीय आयुर्विज्ञान संस्थान, भोपाल
ALL INDIA INSTITUTE OF MEDICAL SCIENCES (AIIMS) BHOPAL

Saket Nagar, Bhopal-462020 (India)
Website : www.aiimsbhopal.edu.in

Pg-10 point no. 10	The Director, AIIMS Bhopal has full authority to take into account the performance of the manufacturer/authorized dealer or distributor/bidder and they should submit a latest performance certificate from any other government Hospitals/institutions/PSUs to testify the proper dealing & performance as well as installation and maintenance of equipment.	Instead of only Indian institutions other institutions world wide should also be considered.	This will give multinational companies to participate with latest technologies	<i>Suggested amendment not accepted</i>
Pg-15 point no. 8	List of name of the organizations and their contact details where the similar nature items previously supplied by him as asked in the schedule of Requirement against this E-tender enquiry by any interested bidder for bidding items to any government Hospitals/Institutions/PSUs/other reputed institutions in Indian in last three financial years(i.e. 2015-2016, 2016-2017, 2017-2018) must required to be uploaded. If any bidder will not upload the requested List of above requested information in respect of items asked in this E-tender Enquiry of the requested FYs, their id is liable for rejection.	Instead of only Indian institutions other institutions world wide should also be considered.	This will give multinational companies to participate with latest technologies	<i>Suggested amendment not accepted</i>

Shashank
01/01/2020

RBK
01-01-2020